



California Medical Device Recall Information



Recall Name

Elite Biomedical Solutions Recalls

Alaris Medley Large Volume Pump Frame Membrane

Due To Potential for Over or Under Delivery of Fluid by Infusion Pumps

Recall Date	Product Description	Recalling Firm	Recall Reason
6/3/15	Alaris Medley Large Volume Pump Frame Membrane	Elite Biomedical Solutions Cincinnati, OH	<i>Potential for over or under delivery of fluid by an infusion pump.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	<u>Frame Membrane NOEM</u> Part Number: <ul style="list-style-type: none">• TC10006587 / 100138041 Lot Numbers: <ul style="list-style-type: none">• 022015502• 0421151000 <u>Frame Membrane NOEM Shipped with LVP Bezel Assembly</u> Part Number: <ul style="list-style-type: none">• 49000204 Lot Numbers: <ul style="list-style-type: none">• 031815100• 032415100• 043015215	CA , nationwide	Manufactured from: February 20 2015 to April 21, 2015 Distributed from: February 25, 2015 to May 8, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm460079.htm>